



RETURN BIDS TO:

**RETOURNER LES
SOUSSIONS À :**

NRC.BidReceiving-
ReceptiondesSoumissions.CNRC@nrc-cnrc.gc.ca

**Amendment #2
Modification #2**

**REQUEST FOR PROPOSAL
DEMANDE DE PROPOSITIONS**

Proposal To: National Research Council Canada

We hereby offer to sell to His Majesty the King in right of Canada, in accordance with the terms and conditions set out herein, referred to herein or attached hereto, the goods, services, and construction listed herein and on any attached sheets at the price(s) set out thereof.

Proposition au : Conseil national de recherches Canada

Nous offrons par la présente de vendre à Sa Majesté le Roi du chef du Canada, aux conditions énoncées ou incluses par référence dans la présente et aux annexes ci-jointes, les biens, services et construction énumérés ici sur toute feuille ci-annexées, au(x) prix indiqué(s).

Instructions : See Herein

Instructions: Voir aux présentes
Comments - Commentaires

**Vendor/Firm Name and address
Raison sociale et adresse du
fournisseur/de l'entrepreneur**

**Issuing Office – Bureau de distribution
National Research Council Canada
Conseil national de recherches Canada**

Title – Sujet monochromator based microplate reader for fluorescence and colorimetric (absorbance) assays	
Solicitation No. – N° de l'invitation 23-58155	Date January 18 th 2024
Solicitation Closes – L'invitation prend fin at – à 02 :00 PM on – le January 23rd 2024	Time Zone Fuseau horaire EST
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Inquiries to : - Adresser toutes questions à: Jonathan Soles	
Email address – l'adresse courriel : Jonathan.Soles@nrc-cnrc.gc.ca	
Destination – of Goods, Services, and Construction: Destination – des biens, services et construction : National Research Council of Canada	
National Research Council of Canada 6100 AVENUE ROYALMOUNT H4P 2R2, MONTRÉAL, Canada	

Vendor/firm Name and address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Facsimile No. – N° de télécopieur Telephone No. – N° de téléphone	
Name and title of person authorized to sign on behalf of Vendor/firm (type or print)- Nom et titre de la personne autorisée à signer au nom du fournisseur/de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date



This amendment is raised to address the following questions and affect the following changes:

Question #2: I am writing to ask for clarification on a point in the "Specifications Required" portion of Annex "A" of RFP 23-58155.

For the following required specification: Must include a Shaker (linear and orbital) and incubator (below 10°C and up to 42°C)

The lower range is dictated by ambient temperature, however. Is this "below 10°C" deliberate and accurate? Is NRC looking for a microplate reader with active cooling?

Answer #2: See the following changes below

Change #1:

DELETED: "• Must include a Shaker (linear and orbital) and incubator (below 10°C and up to 42°C)" IN ANNEX A

INSERT:

Must include a Shaker (linear and orbital) and incubator (Operating range must be from ambient temperature +5°C to 42°C)." IN ANNEX A

Change #2 :

DELETE: "• Must include a Shaker (linear and orbital) and incubator (below 10°C and up to 42°C)" IN ANNEX C Mandatory Technical Evaluation Criteria

INSERT: the following row at the bottom of Annex C's table of Mandatory technical Evaluation criteria.

	<input type="checkbox"/> Must include a Shaker (linear and orbital) and incubator (Operating range must be from ambient temperature +5°C to 42°C).	Document Name: _____ Page Number: _____	YES/NO
--	--	--	--------