



Royal Canadian Gendarmerie royale
Mounted Police du Canada

RETURN BIDS TO:
RETOURNER LES OFFRES A:

ATTENTION: Shannon Plunkett
Email: Shannon.Plunkett@rcmp-grc.gc.ca

REQUEST FOR PROPOSAL AMENDMENT

MODIFICATION DE DEMANDE DE PROPOSITION

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments: - Commentaries :

Title – Sujet Nitrogen Generators		Date May 7th 2024
Solicitation No. – N° de l'invitation 202401449/A		Amendment No. – N° de la modification 001
Client Reference No. - No. De Référence du Client 202401449		
Solicitation Closes – L'invitation prend fin		
At / à :	02 :00 PM	EST(EasternStandard Time) HNE (heure normale de l'Est)
On / le :	May 10th, 2024	
Delivery - Livraison See herein — Voir aux présentes	Taxes - Taxes See herein — Voir aux présentes	Duty – Droits See herein — Voir aux présentes
Destination of Goods and Services – Destinations des biens et services See herein — Voir aux présentes		
Instructions See herein — Voir aux présentes		
Address Inquiries to – Adresser toute demande de renseignements à Shannon Plunkett Shannon.Plunkett@rcmp-grc.gc.ca		
Telephone No. – No. de téléphone 613-794-2264		

Delivery Required – Livraison exigée See herein — Voir aux présentes	Delivery Offered – Livraison proposée
Vendor/Firm Name, Address and Representative – Raison sociale, adresse et représentant du fournisseur/de l'entrepreneur:	
Telephone No. – No. de téléphone	Facsimile No. – No. de télécopieur
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) – Nom et titre de la personne autorisée à signer au nom du fournisseur/de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date



Amendment 001 is raised to address the following Questions

Q: In response to subject solicitation, would you please extend the due date for two weeks to allow all interested suppliers prepare a proposal?

A: Because we are trying to address health and safety concerns due to heat and noise, we are not willing to extend for another two weeks.

Q: Can you please advise on the questions below:
In reference to ANNEX A - STATEMENT OF REQUIREMENT, 5.3.2 Installation:

Q1: Is it just a connection to the air line or a complete set up?

A: Q1: Partial set up. there is already a line from the air compressor into the ceiling, so the units would need to be connected to that line, which is currently being used for another nitrogen generator. So a T would likely need to be installed

Q2: If that is the case, how far are the air lines?

A: Q2: Current line is probably less than 5 metres away from the wall, possibly less

Q3: How far are the air lines to the end point?

A: Q3: In addition to the 5 m or so in the ceiling, the drop to the ceiling to the units would likely be approximately 2 m

Q4: Is the endpoint in the same room?

A: Q4: The end point would be in the same room as the current line

Q5: Is there one or more end points?

In reference to ANNEX A - STATEMENT OF REQUIREMENT, 5.1.1 "Each proposed nitrogen generator must not exceed 50 pounds (lbs) in weight.":

A: Q5: Only one end point

Q6: Is a product with a weight of < 50 kg (110 pounds) acceptable?

In reference to ANNEX A - STATEMENT OF REQUIREMENT, 5.1.1 "The nitrogen generator(s) must not require electricity (no requirements for electrical outlets).":

A: Q6: If a single generator can provide the requirements, this would be acceptable

Q7: Is a product with a supply rating of 100-240 Vac ($\pm 10\%$) 50/60Hz acceptable?

A: Q7: We require no electrical requirements, as this is contributing to the heat generation and overheating of the room. so, a supply rating of 100-240 VAC would not be acceptable.



Royal Canadian Mounted Police Gendarmerie royale du Canada

Solicitation No. – N° de l'invitation : 202401449/A
Amendment No. – No. de la modification : 001

Delete:

Solicitation Closes May 7th, 2024 at 2PM EDT

Insert:

Solicitation Closes May 10th, 2024 at 2PM EDT