

Amendment No. 002

This amendment is being raised to provide a response to Request for Clarification.

Question 1.

Q1. Annex A, Item 1a. Is a system that provides 5V maximum output voltage acceptable?

R1. Yes, it is acceptable.

Question 2.

Q2. Annex A, Item 1b. Is a system that provides 100A maximum output current acceptable?

R2. Yes, it is acceptable.

Question 3.

Q3. Annex A, Item 4b. Is a maximum pressure for liquid side 20 PSI acceptable?

R3. No, it is not acceptable. The pressure for the back pressure control must be operated in the range of 0 psig to 30 psig, or higher.

Question 4.

Q4. Annex A, Item 5f. The requested system does not have liquid level monitoring for each reservoir. Will Canada consider removing this requirement from RFP?

R4. No, it is not acceptable. The unit must have liquid level monitoring for each reservoir.

All other terms and conditions of the solicitation remain the same.